

EKOTEKS LABORATUVAR ve GÖZETİM  
HİZMETLERİ A.Ş.


AB-0583-T

20038512-  
ing

10-20

REQUIRED TESTS	RESULT	COMMENTS
<b>MICROBIOLOGICAL TESTS</b>		
Bacterial Filtration Efficiency / <b>Original</b>	P	
Bacterial Filtration Efficiency / <b>After 5 Washes</b>	P	
Microbial Cleanliness(Bioburden) / <b>Original</b>	P	
Microbial Cleanliness(Bioburden) / <b>After 5 Washes</b>	P	
<b>PHYSICAL PROPERTIES</b>		
Breathability(Differential Pressure) / <b>Original</b>	P	
Breathability(Differential Pressure) / <b>After 5 Washes</b>	P	
P: Pass F: Fail R: Refer to retailer technologist. Test results were evaluated according to TSE K 599: 2020 limit values		

REMARK: Original samples are kept for 3 months and all technical records are kept for 5 years unless otherwise specified. If requested, measurement uncertainty will be reported. But unless otherwise specified, measurement uncertainty is not considered while stating compliance with specification or limit values. The reported uncertainty is based on a standard uncertainty multiplied by a coverage factor  $k=2$ , providing a level of confidence of approximately 95 %. Tests marked (\*) in this report are not included in the accreditation schedule.



Gen.f136-2/03

This report shall not be reproduced other than in full except with the permission of the laboratory.  
*Testing reports without signature and seal are not valid.*